



Human VEGF-A (Vascular Endothelial Cell Growth Factor A) ELISA Kit

Cat# E5053

INFORMATION

Catalog Number	E5053
Size	96T
Sensitivity	18.75 pg/mL
Detection Range	31.25-2000 pg/mL
Specificity	This kit recognizes Human VEGF-A in samples. No significant cross-reactivity or interference between Human VEGF-A and analogues was observed.
Repeatability	Coefficient of variation is < 10%
Test principle	This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human VEGF-A. Samples (or Standards) are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human VEGF-A and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human VEGF-A, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm \pm 2 nm. The OD value is proportional to the concentration of Human VEGF-A. You can calculate the concentration of Human VEGF-A in the samples by comparing the OD of the samples to the standard curve.
Other supplies required	Microplate reader with 450 nm wavelength filter High-precision transfer pipette, EP tubes and disposable pipette tips Incubator capable of maintaining 37°C Deionized or distilled water Absorbent paper Loading slot
Sample collection	Serum: Allow samples to clot for 1 hour at room temperature or overnight at 2-8°C before centrifugation for 20 min at 1000×g at 2-8°C. Collect the supernatant to carry out the assay. Plasma: Collect plasma using EDTA-Na ₂ as an anticoagulant. Centrifuge samples for 15 min at 1000×g at 2-8°C within 30 min of collection. Collect the supernatant to carry out the assay. Tissue homogenates: It is recommended to get detailed references from the literature before analyzing different tissue types. For general information, hemolyzed blood may affect the results, so the tissues should be minced into small pieces and rinsed in

	<p>ice-cold PBS (0.01M, pH=7.4) to remove excess blood thoroughly. Tissue pieces should be weighed and then homogenized in PBS (tissue weight (g): PBS (mL) volume=1:9) with a glass homogenizer on ice. To further break down the cells, you can sonicate the suspension with an ultrasonic cell disrupter or subject it to freeze-thaw cycles. The homogenates are then centrifuged for 5-10 min at 5000×g at 2-8°C to get the supernatant.</p> <p>Cell lysates: For adherent cells, gently wash the cells with moderate amount of pre-cooled PBS and dissociate the cells using trypsin. Collect the cell suspension into a centrifuge tube and centrifuge for 5 min at 1000×g. Discard the medium and wash the cells 3 times with pre-cooled PBS. For each 1×10⁶ cells, add 150-250 μL of pre-cooled PBS to keep the cells suspended. Repeat the freeze-thaw process several times or use an ultrasonic cell disrupter until the cells are fully lysed. Centrifuge for 10 min at 1500×g at 2-8°C. Remove the cell fragments, collect the supernatant to carry out the assay.</p> <p>Cell culture supernatant or other biological fluids: Centrifuge samples for 20 min at 1000×g at 2-8°C. Collect the supernatant to carry out the assay.</p>
<p>Note for kit</p>	<ol style="list-style-type: none"> 1) For research use only. Not for use in diagnostic procedures. 2) Please wear lab coats, eye protection and latex gloves for protection. Please perform the experiment following the national security protocols of biological laboratories, especially when detecting blood samples or other bodily fluids. 3) A freshly opened ELISA plate may appear a water-like substance, which is normal and will not have any impact on the experimental results. Return the unused wells to the foil pouch and store according to the conditions suggested in the above table. 4) Do not reuse the reconstituted standard, biotinylated detection Ab working solution, concentrated HRP conjugate working solution. The unspent undiluted concentrated biotinylated detection Ab (100×) and other stock solutions should be stored according to the storage conditions in the above table. 5) The microplate reader should be able to be installed with a filter that can detect the wave length at 450±10 nm. The optical density should be within 0-3.5. Follow the Instructions of the Microplate Reader for set-up and preheat it for 15 min before OD measurement. 6) Do not mix or substitute reagents with those from other lots or sources. 7) Change pipette tips in between adding of each standard level, between sample adding and between reagent adding. Also, use separate reservoirs for each reagent. 8) The kit should not be used beyond the expiration date on the kit label.
<p>Note for sample</p>	<ol style="list-style-type: none"> 1) Tubes for blood collection should be disposable and be non-endotoxin. Samples with high hemolysis or much lipid are not suitable for ELISA assay. 2) Samples should be assayed within 7 days when stored at 2-8°C,

	<p>otherwise samples must be divided up and stored at -20°C (≤1 month) or -80°C (≤3 months). Avoid repeated freeze-thaw cycles. Prior to assay, the frozen samples should be slowly thawed and centrifuged to remove precipitates.</p> <p>3) Please predict the concentration before assaying. If the sample concentration is not within the range of the standard curve, users must determine the optimal sample dilutions for their particular experiments.</p> <p>4) If the sample type is not included in the manual, a preliminary experiment is suggested to verify the validity.</p> <p>5) If a lysis buffer is used to prepare tissue homogenates or cell lysates, there is a possibility of causing a deviation due to the introduced chemical substance.</p> <p>6) Some recombinant protein may not be detected due to a mismatching with the coated antibody or detection antibody.</p>
Dilution Method	<p>Please predict the concentration range of the sample in advance. If your test sample needs dilution, please refer to the dilution method as follows: For 100 fold dilution: One-step dilution. Add 5 µL sample to 495 µL sample diluent to yield 100 fold dilution. For 1000 fold dilution: Two-step dilution. Add 5 µL sample to 95 µL sample diluent to yield 20 fold dilution, then add 5 µL 20 fold diluted sample to 245 µL sample diluent, after this, the neat sample has been diluted at 1000 fold successfully. For 100000 fold dilution: Three-step dilution. Add 5 µL sample to 195 µL sample diluent to yield 40 fold dilution, then add 5 µL 40 fold diluted sample to 245 µL sample diluent to yield 50 fold dilution, and finally add 5 µL 2000 fold diluted sample to 245 µL sample diluent, after this, the neat sample has been diluted at 100000 fold successfully.</p>
Reagent preparation	<p>1. Bring all reagents to room temperature (18-25°C) before use. If the kit will not be used up in one assay, please only take out the necessary strips and reagents for present experiment, and store the remaining strips and reagents at required condition.</p> <p>2. Wash Buffer: Dilute 30 mL of Concentrated Wash Buffer with 720 mL of deionized or distilled water to prepare 750 mL of Wash Buffer. Note: if crystals have formed in the concentrate, warm it in a 40°C water bath and mix it gently until the crystals have completely dissolved.</p> <p>3. Standard working solution: Centrifuge the standard at 10,000×g for 1 min. Add 1.0 mL of Reference Standard&Sample Diluent, let it stand for 10 min and invert it gently several times. After it dissolves fully, mix it thoroughly with a pipette. This reconstitution produces a working solution of 2000 pg/mL(or add 1 mL of Reference Standard&Sample Diluent, let it stand for 1-2 min and then mix it thoroughly with a vortex meter of low speed. Bubbles generated during vortex could be removed by centrifuging at a relatively low speed). Then make serial dilutions as needed. The recommended dilution gradient is as follows: 2000, 1000, 500, 250, 125, 62.5, 31.25, 0 pg/mL. Dilution method: Take 7 EP tubes, add 500uL of Reference Standard & Sample Diluent to each tube. Pipette 500uL of the 2000 pg/mL working solution to the first tube and mix up to produce a 1000 pg/mL working solution. Pipette 500uL of the solution from the former tube into the latter one according to this step. The illustration below is for reference.</p> <p>4. Biotinylated Detection Ab working solution: Calculate the required amount</p>

	<p>before the experiment (100 μL/well).In preparation, slightly more than calculated should be prepared. Centrifuge the Concentrated Biotinylated Detection Ab at 800\timesg for 1 min, then dilute the 100\times Concentrated Biotinylated Detection Ab to 1\times working solution with Biotinylated Detection Ab Diluent(Concentrated Biotinylated Detection Ab: Biotinylated Detection Ab Diluent= 1:99).</p> <p>5. Concentrated HRP Conjugate working solution: Calculate the required amount before the experiment (100 μL/well). In preparation, slightly more than calculated should be prepared. Centrifuge the Concentrated HRP Conjugate at 800\timesg for 1 min, then dilute the 100\times Concentrated HRP Conjugate to 1\times working solution with HRP Conjugate Diluent(Concentrated HRP Conjugate: HRP Conjugate Diluent= 1: 99).</p>
<p>Assay procedure</p>	<ol style="list-style-type: none"> 1. Determine wells for diluted standard, blank and sample. Add 100 μL each dilution of standard, blank and sample into the appropriate wells (It is recommended that all samples and standards be assayed in duplicate). Cover the plate with the sealer provided in the kit. Incubate for 90 min at 37$^{\circ}$C. Note: solutions should be added to the bottom of the micro ELISA plate well, avoid touching the inside wall and causing foaming as much as possible. 2. Decant the liquid from each well, do not wash. Immediately add 100 μL of Biotinylated Detection Ab working solution to each well. Cover the plate with a new sealer. Incubate for 1 hour at 37$^{\circ}$C. 3. Decant the solution from each well, add 350 μL of wash buffer to each well. Soak for 1 min and aspirate or decant the solution from each well and pat it dry against clean absorbent paper. Repeat this wash step 3 times. Note: a microplate washer can be used in this step and other wash steps. Make the tested strips in use immediately after the wash step. Do not allow wells to be dry. 4. Add 100 μL of HRP Conjugate working solution to each well. Cover the plate with a new sealer. Incubate for 30 min at 37$^{\circ}$C. 5. Decant the solution from each well, repeat the wash process for 5 times as conducted in step 3. 6. Add 90 μL of Substrate Reagent to each well. Cover the plate with a new sealer. Incubate for about 15 min at 37$^{\circ}$C. Protect the plate from light. Note: the reaction time can be shortened or extended according to the actual color change, but not more than 30 min. Preheat the Microplate Reader for about 15 min before OD measurement. 7. Add 50 μL of Stop Solution to each well. Note: adding the stop solution should be done in the same order as the substrate solution. 8. Determine the optical density (OD value) of each well at once with a micro-plate reader set to 450 nm.
<p>Calculation of results</p>	<p>Average the duplicate readings for each standard and samples, then subtract the average zero standard optical density. Plot a four parameter logistic curve on log-log axis, with standard concentration on the x-axis and OD values on the yaxis. If the OD of the sample surpasses the upper limit of the standard curve, you should re-test it with an appropriate dilution.</p>

	The actual concentration is the calculated concentration multiplied by the dilution factor.																																									
Typical data	<p>As the OD values of the standard curve may vary according to the conditions of the actual assay performance (e.g.operator, pipetting technique, washing technique or temperature effects), the operator should establish a standard curve for each test. Typical standard curve and data is provided below for reference only.</p> <table border="1"> <thead> <tr> <th>pg/mL</th> <th>OD</th> <th>Corrected OD</th> <th rowspan="9"> </th> </tr> </thead> <tbody> <tr> <td>2000</td> <td>2.496</td> <td>2.412</td> </tr> <tr> <td>1000</td> <td>1.615</td> <td>1.531</td> </tr> <tr> <td>500</td> <td>0.923</td> <td>0.839</td> </tr> <tr> <td>250</td> <td>0.444</td> <td>0.36</td> </tr> <tr> <td>125</td> <td>0.267</td> <td>0.183</td> </tr> <tr> <td>62.5</td> <td>0.184</td> <td>0.1</td> </tr> <tr> <td>31.25</td> <td>0.136</td> <td>0.052</td> </tr> <tr> <td>0</td> <td>0.084</td> <td>-</td> </tr> </tbody> </table>	pg/mL	OD	Corrected OD		2000	2.496	2.412	1000	1.615	1.531	500	0.923	0.839	250	0.444	0.36	125	0.267	0.183	62.5	0.184	0.1	31.25	0.136	0.052	0	0.084	-													
pg/mL	OD	Corrected OD																																								
2000	2.496	2.412																																								
1000	1.615	1.531																																								
500	0.923	0.839																																								
250	0.444	0.36																																								
125	0.267	0.183																																								
62.5	0.184	0.1																																								
31.25	0.136	0.052																																								
0	0.084	-																																								
Precision	<p>Intra-assay Precision (Precision within an assay): 3 samples with low, mid range and high level Human VEGF-A were tested 20 times on one plate, respectively.</p> <p>Inter-assay Precision (Precision between assays): 3 samples with low, mid range and high level Human VEGF-A were tested on 3 different plates, 20 replicates in each plate, respectively.</p> <table border="1"> <thead> <tr> <th rowspan="2">Sample</th> <th colspan="3">Intra-assay Precision</th> <th colspan="3">Inter-assay Precision</th> </tr> <tr> <th>1</th> <th>2</th> <th>3</th> <th>1</th> <th>2</th> <th>3</th> </tr> </thead> <tbody> <tr> <td>n</td> <td>20</td> <td>20</td> <td>20</td> <td>20</td> <td>20</td> <td>20</td> </tr> <tr> <td>Mean(pg/mL)</td> <td>94.10</td> <td>279.80</td> <td>737.20</td> <td>92.40</td> <td>303.90</td> <td>745.00</td> </tr> <tr> <td>Standard deviation</td> <td>5.50</td> <td>14.30</td> <td>34.60</td> <td>4.80</td> <td>15.50</td> <td>31.30</td> </tr> <tr> <td>C V (%)</td> <td>5.84</td> <td>5.11</td> <td>4.69</td> <td>5.19</td> <td>5.10</td> <td>4.20</td> </tr> </tbody> </table>	Sample	Intra-assay Precision			Inter-assay Precision			1	2	3	1	2	3	n	20	20	20	20	20	20	Mean(pg/mL)	94.10	279.80	737.20	92.40	303.90	745.00	Standard deviation	5.50	14.30	34.60	4.80	15.50	31.30	C V (%)	5.84	5.11	4.69	5.19	5.10	4.20
Sample	Intra-assay Precision			Inter-assay Precision																																						
	1	2	3	1	2	3																																				
n	20	20	20	20	20	20																																				
Mean(pg/mL)	94.10	279.80	737.20	92.40	303.90	745.00																																				
Standard deviation	5.50	14.30	34.60	4.80	15.50	31.30																																				
C V (%)	5.84	5.11	4.69	5.19	5.10	4.20																																				
Recovery	<p>The recovery of Human VEGF-A spiked at three different levels in samples throughout the range of the assay was evaluated in various matrices.</p> <table border="1"> <thead> <tr> <th>Sample Type</th> <th>Range (%)</th> <th>Average Recovery (%)</th> </tr> </thead> <tbody> <tr> <td>Serum (n=8)</td> <td>92-105</td> <td>98</td> </tr> <tr> <td>EDTA plasma (n=8)</td> <td>95-110</td> <td>101</td> </tr> <tr> <td>Cell culture media (n=8)</td> <td>88-103</td> <td>96</td> </tr> </tbody> </table>	Sample Type	Range (%)	Average Recovery (%)	Serum (n=8)	92-105	98	EDTA plasma (n=8)	95-110	101	Cell culture media (n=8)	88-103	96																													
Sample Type	Range (%)	Average Recovery (%)																																								
Serum (n=8)	92-105	98																																								
EDTA plasma (n=8)	95-110	101																																								
Cell culture media (n=8)	88-103	96																																								
Linearity	<p>Samples were spiked with high concentrations of Human VEGF-A and diluted with Reference Standard & Sample Diluent to produce samples with values within the range of the assay.</p> <table border="1"> <thead> <tr> <th rowspan="2"></th> <th rowspan="2"></th> <th>Serum (n=5)</th> <th>EDTA plasma(n=5)</th> <th>Cell culture media(n=5)</th> </tr> </thead> <tbody> <tr> <td>1:2</td> <td>Range (%) 99-114 Average (%) 105</td> <td>97-111 104</td> <td>91-105 99</td> </tr> <tr> <td>1:4</td> <td>Range (%) 95-112 Average (%) 103</td> <td>80-95 87</td> <td>93-106 98</td> </tr> <tr> <td>1:8</td> <td>Range (%) 100-116 Average (%) 106</td> <td>86-100 93</td> <td>97-110 103</td> </tr> <tr> <td>1:16</td> <td>Range (%) 96-110 Average (%) 104</td> <td>82-92 87</td> <td>96-107 102</td> </tr> </tbody> </table>			Serum (n=5)	EDTA plasma(n=5)	Cell culture media(n=5)	1:2	Range (%) 99-114 Average (%) 105	97-111 104	91-105 99	1:4	Range (%) 95-112 Average (%) 103	80-95 87	93-106 98	1:8	Range (%) 100-116 Average (%) 106	86-100 93	97-110 103	1:16	Range (%) 96-110 Average (%) 104	82-92 87	96-107 102																				
				Serum (n=5)	EDTA plasma(n=5)	Cell culture media(n=5)																																				
		1:2	Range (%) 99-114 Average (%) 105	97-111 104	91-105 99																																					
1:4	Range (%) 95-112 Average (%) 103	80-95 87	93-106 98																																							
1:8	Range (%) 100-116 Average (%) 106	86-100 93	97-110 103																																							
1:16	Range (%) 96-110 Average (%) 104	82-92 87	96-107 102																																							
Declaration	1. Limited by current conditions and scientific technology, we can't conduct																																									

comprehensive identification and analysis on all the raw material provided. So there might be some qualitative and technical risks for users using the kit.

2. This assay is designed to eliminate interference by factors present in biological samples. Until all factors have been tested in the ELISA immunoassay, the possibility of interference cannot be excluded.

3. The final experimental results will be closely related to the validity of products, operational skills of the operators, the experimental environments and so on. We are only responsible for the kit itself, but not for the samples consumed during the assay. The users should calculate the possible amount of the samples used in the whole test. Please reserve sufficient samples in advance.

4. To get the best results, please only use the reagents supplied by the manufacturer and strictly comply with the instructions.

5. Incorrect results may occur because of incorrect operations during the reagents preparation and loading, as well as incorrect parameter settings of the Micro-plate reader. Please read the instructions carefully and adjust the instrument prior to the experiment.

6. Even the same operator might get different results in two separate experiments. In order to get reproducible results, the operation of every step in the assay should be controlled.

7. Every kit has strictly passed QC test. However, results from end users might be inconsistent with our data due to some variables such as transportation conditions, different lab equipment, and so on. Intra-assay variance among kits from different batches might arise from the above reasons too.

8. Kits from different manufacturers or other methods for testing the same analyte could bring out inconsistent results, since we haven't compared our products with those from other manufacturers.

9. The kit is designed for research use only, we will not be responsible for any issues if the kit is applied in clinical diagnosis or any other related procedures.

Troubleshooting

Problem	Causes	Solutions
Poor standard curve	Inaccurate pipetting	Check pipettes.
	Improper standard dilution	Ensure briefly spin the vial of standard and dissolve the powder thoroughly by gentle mixing.
	Wells are not completely aspirated	Completely aspirate wells in between steps.
Low signal	Insufficient incubation time	Ensure sufficient incubation time.
	Incorrect assay temperature	Use recommended incubation temperature. Bring substrate to room temperature before use.
	Inadequate reagent volumes	Check pipettes and ensure correct preparation.
	Improper dilution	
HRP conjugate inactive or TMB failure	Mix HRP conjugate and TMB, rapid coloring.	
Deep color but low value	Plate reader setting is not optimal	Verify the wavelength and filter setting on the Microplate reader.
		Open the Microplate Reader ahead to pre-heat.
Large CV	Inaccurate pipetting	Check pipettes.
High background	Concentration of target protein is too high	Use recommended dilution factor.
	Plate is insufficiently washed	Review the manual for proper wash. If using a plate washer, check that all ports are unobstructed.
	Contaminated wash buffer	Prepare fresh wash buffer.
Low sensitivity	Improper storage of the ELISA kit	All the reagents should be stored according to the instructions.
	Stop solution is not added	Stop solution should be added to each well before measurement.